



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/663,488	09/16/2003	Joseph P. Errico	F-291	2425

530 7590 10/18/2006

LERNER, DAVID, LITTENBERG,
KRUMHOLZ & MENTLIK
600 SOUTH AVENUE WEST
WESTFIELD, NJ 07090

EXAMINER

PELLEGRINO, BRIAN E

ART UNIT	PAPER NUMBER
----------	--------------

3738

DATE MAILED: 10/18/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/663,488

Applicant(s)

ERRICO ET AL.

Examiner

Brian E. Pellegrino

Art Unit

3738

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 07 August 2006.
 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-14 is/are pending in the application.
 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
 5) ☐ Claim(s) _____ is/are allowed.
 6) ☒ Claim(s) 1-14 is/are rejected.
 7) ☐ Claim(s) _____ is/are objected to.
 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 a) ☐ All b) ☐ Some * c) ☐ None of:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
 Paper No(s)/Mail Date _____.
 4) ☐ Interview Summary (PTO-413)
 Paper No(s)/Mail Date. _____.
 5) ☐ Notice of Informal Patent Application (PTO-152)
 6) ☐ Other: _____.

DETAILED ACTION

Claim Rejections - 35 USC § 102

The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

Claims 1-14 are rejected under 35 U.S.C. 102(b) as being anticipated by Young et al. (6190414). Fig. 10 shows an intervertebral spacer device with first and second baseplates (**158,160**) that are mounted to one another via joint **162**. It can also be seen the baseplates include a plurality of engagement holes (**170,172**) in a perimetrical region that has a single post **140** positioned within the holes of the plates. The examiner is interpreting the claimed element "engagement hole" in this way: a cavity in something solid or an opening. Claims in a pending application should be given their broadest reasonable interpretation. *In re Pearson*, 181 USPQ 641 (CCPA 1974). See also *In re Morris*, Fed. Cir. 1997 127 F3d 1048, 1054,1055. Please note the intended use, as set forth in the claims, carries no weight in the absence of any distinguishing structure. Thus the spinal device is fully capable of having the desired surgical approach aspects. It can be seen that the post has a first position with a significant portion of the post within the tool and when cranked up to engage the plates to a second position, the post extends outward from the distal end to dispose within the engagement holes of the plates.

Response to Arguments

Applicant's arguments filed 8/7/06 have been fully considered but they are not persuasive. Applicant argues that Young's device does not have the post within the tool. However, the Examiner would like to refer Applicants to Figs. 6 & 7 that illustrate that a significant portion of the post is clearly positioned within the tool before jacking up the plates of the implant. During the procedure the post is then jacked to extend outward in a second final position to engage the plates. Applicant also argues that only a first end of the post engages the engagement hole of the plate and Young does not show this. It is the Examiner's position that Young does indeed disclose this limitation because the post has one end engaging the plate and the other end is within a gear cylinder.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. EP (122903) discloses a tool with a slidable pin to insert a prosthesis having two plates.

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within

TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brian E. Pellegrino whose telephone number is 571-272-4756. The examiner can normally be reached on Monday-Thursday from 7am to 4:30pm. The examiner can also be reached on alternate Fridays. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott, can be reached at 571-272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

TC 3700, AU 3738

BRIAN E. PELLEGRINO
PRIMARY EXAMINER
